Methods

Fluorescence Polarization in Homogeneous Nucleic Acid Analysis II: 5'-Nuclease Assay

Sherif Latif,^{1,4} Irma Bauer-Sardina,^{1,4} Kostubh Ranade,² Kenneth J. Livak,³ and Pui-Yan Kwok^{1,5}

¹*Division of Dermatology, Washington University School of Medicine, St. Louis, Missouri 63110, USA;* ²*Department of Genetics, Stanford University School of Medicine, Stanford, California 94305, USA;* ³*Applied Biosystems, Foster City, California 94404, USA*

When the temperature and viscosity of the solvent is held constant, the degree of fluorescence polarization (FP) detected when a fluorescent dye is excited by plane polarized light depends mostly on the molecular weight of the dye molecule. By monitoring the FP of a fluorescent dye molecule, one can detect significant changes in the molecular weight of a fluorescent molecule without separation or purification. The 5'-nuclease (TaqMan) assay is a robust single nucleotide polymorphism genotyping method where an allele-specific probe that binds to a perfectly complementary target is cleaved by the 5'-nuclease activity of Taq DNA polymerase. Because the TaqMan probe is labeled with a fluorescent dye, it has high FP value when intact but a low FP value after cleavage. In this study, we compared the results of the 5'-nuclease assay based on standard fluorescence intensity readings and FP readings when genotyping 90 individuals with 20 single nucleotide polymorphisms. Our results show that FP is just as robust and reliable as the standard fluorescence detection method. Use of FP detection makes it possible to reduce the cost of TaqMan probes by abrogating the need for a fluorescence quencher.

The ability to determine efficiently and unambiguously the mutational status or genotype of an organism has great applications in molecular diagnostics, clinical genetic testing, population genetics, and agricultural biotechnology. High-throughput genotyping methods for single nucleotide variations currently in use discriminate between the alleles by differential hybridization, primer extension, ligation, and allele-specific cleavage of a probe (Kwok 2000). Homogeneous assays based on these allele discrimination mechanisms are conducted in aqueous solutions without separation or purification by monitoring physical changes when the reagents are turned into products. We have shown previously that fluorescence polarization (FP) is a good detection method in the primer extension assay when a dye-labeled dideoxy terminator is incorporated allele-specifically in the presence of a matching DNA template (Chen et al. 1999). In this report, we show that FP also is a good detection method for the 5'-nuclease (TaqMan) assay, where a fluorescent probe is cleaved during the polymerase chain reaction only when it is annealed to a perfectly complementary template.

When a fluorescent molecule is excited by plane-polarized light at the correct wavelength, the fluorescence emitted is also polarized. However, because the

molecule rotates and tumbles in space, the FP observed is proportional to the fluorescent molecule's rotational relaxation time (the time it takes to rotate through an angle of 68.5°), which is related to the viscosity of the solvent, absolute temperature, molecular volume, and the gas constant. Therefore, if the viscosity and temperature are constant, FP is directly proportional to the molecular volume, which is directly proportional to molecular weight. If the fluorescent molecule is large (with high molecular weight), it rotates and tumbles more slowly in space and FP is preserved. If the molecule is small (with low molecular weight), it rotates and tumbles faster and FP is largely lost or depolarized. In principle, FP can be used to detect any significant change in molecular weight of a fluorescent molecule. Indeed, FP detection is the basis of numerous clinical and research assays, especially those involved in ligand-receptor binding (Checovich et al. 1995).

The 5'-nuclease assay is one of the simplest diagnostic assays by which one can determine the mutational status of a DNA sample in one step (Livak 1999). The current detection method in the 5'-nuclease assay relies on the increase in fluorescence intensity when a reporter fluorophore is released from its quencher as the doubly labeled probe is cleaved during PCR (Lee et al. 1999). Because the 5'-nuclease assay is one where a large probe is cleaved into small molecules, we reason that FP can be a good detection system for the method (Fig. 1).

To show that FP detection is indeed suitable for the 5'-nuclease assay, we performed the assay with 20

⁴**These authors contributed equally to this work.**

⁵**Corresponding author.**

E-MAIL kwok@genetics.wustl.edu; **FAX** (314) 362-8159.

Article and publication are at www.genome.org/cgi/doi/10.1101/gr.156601

Figure 1 5'-nuclease assay with FP detection.

markers on DNA samples from 90 individuals. At the end of the assay, detection and analysis was done by both fluorescence intensity and FP. We report here that both fluorescence intensity and FP gave completely concordant genotypes.

RESULTS

Twenty markers, previously optimized for the 5'-nuclease assay, were used in this study (Table 1). The two allele-specific probes were labeled with reporter dyes VIC (allele 1) and FAM (allele 2), respectively, and contain a nonfluorescent quencher. In addition, the probes also are fitted with a minor groove binder to keep the probe length to a minimum and maximize the effect of the one-base mismatch (Kutyavin et al. 1997; Afonina et al. 1997). Genomic DNA samples of 90 individuals selected from the National Institutes of Health (NIH) Polymorphism Discovery Panel (Collins et al. 1998) were typed for each marker. Six water blanks served as the negative controls. For all 20 sets of experiments (1920 independent genotyping experiments), the fluorescence intensity readings and FP readings at the end of the experiments clustered into four distinct groups (Fig. 2), representing the four possible outcomes of each genotyping experiment. Figure 2a shows the normalized fluorescence intensities of the individual samples at the end of the assay for marker D2S1301, an anonymous single nucleotide polymorphism (SNP) marker developed in our laboratory (Kwok et al. 1996). For the negative controls, both probes remain intact at the end of the reaction and the fluorescence of the reporters is quenched. The data points for the negative controls therefore occupy the region near the origin of the scatter plot, where the fluorescence intensities of both VIC and FAM are low. For the samples homozygous for allele 1, the VIC probe is cleaved but the FAM probe remains intact. They are found in the lower right corner of the plot, where VIC fluorescence is high but FAM fluorescence is low. Similarly, homozygotes with allele 2 occupy the upper left

corner of the plot, where FAM fluorescence is high but VIC fluorescence is low. For the heterozygotes, both probes are cleaved and they occupy the upper right corner of the plot, where both VIC and FAM fluorescence are high.

Figure 2b shows the results of the FP analysis of marker D2S1301. The positive reactions in this case give low FP values (as opposed to high fluorescence intensity values). For the negative controls, both probes remain intact at the end of the reaction and the FP values of both reporters remain high. Therefore, the data points for the negative controls occupy the upper right corner of the scatter plot. For the samples homozygous for allele 1, the VIC probe is cleaved but the FAM probe remains intact. They are found in the lower right corner of the plot, where VIC FP is low but FAM FP is high. Similarly, homozygotes with allele 2 occupy the upper left corner of the plot, where FAM FP is low but VIC FP is high. For the heterozygotes, both probes are cleaved and they occupy the origin of the plot, where both VIC and FAM FP values are low.

Semi-automated genotype assignments are made by taking the ratio between the normalized FP values of VIC and FAM. A sample with a low VIC/FAM ratio represents a homozygous allele 1 sample (cleaved VIC probe, intact FAM probe); a high VIC/FAM ratio represents a homozygous allele 2 sample (intact VIC probe, cleaved FAM probe); a VIC/FAM ratio close to one represents a heterozygote if the absolute FP values of both VIC and FAM are low (both probes are cleaved) but a negative control (or failed PCR) if the absolute FP values of both dyes are high (both probes are intact). Table 2 shows the genotype assignments of 24 DNA samples and six negative controls for marker D2S1301. All genotypes generated by this study have been deposited in dbSNP.

The genotypes called, based on fluorescence intensity readings and FP readings, are compared for all 20 markers (1800 samples total). Table 3 shows the concordance rate of these experiments. All 20 sets of experiments worked the first time, with only three samples where the genotypes could not be called with high confidence (because of weak PCR amplification). When the genotyping assays were redone for the three failed samples, all gave unambiguous and concordant genotypes. In this study, the concordance rate between the fluorescence intensity analysis and FP analysis of the 5'-nuclease assay was 100%.

DISCUSSION

The 5'-nuclease assay is a proven method for the detection of SNPs. In our hands, the 20 markers performed beautifully the first time under a universal set of conditions. In the current configuration, a modified probe containing a fluorescent reporter, a quencher, and a minor groove binder is required for each allele.

Table 1. PCR Primer and TaqMan Probe Sequences for SNP Markers Studied

Marker	NCBI assay ID	Allele		Allele 1 (VIC) TaqMan probe	Allele 2 (FAM) TaqMan probe	PCR Primer 1	PCR Primer 2
		1	2				
D4S2341	231	C	G	VIC-ccatgGgttagca-NFQ-MGB	FAM-ccatgGgttagcc-NFQ-MGB	GAGTTTCACCTGGCTTGCT	CATGACATTATCTAATGATTTTCTTATT
D10S1188	241	A	C	VIC-tgtaccacTcttttc-NFQ-MGB	FAM-taccacGcttttc-NFQ-MGB	TGCTGAATTTCCATCTCTGATTC	GCAGATTTCAGTCCAGAAAG
D2S1301	221	A	G	VIC-ttcatctccAttgtg-NFQ-MGB	FAM-ttcatctccCttgtg-NFQ-MGB	CCCAGTATCTGGCACATC	TTTCTCAGGTTTCAGGGATTACA
D11S1964	245	del	T	VIC-caccccaatagca-NFQ-MGB	FAM-caccccTaataga-NFQ-MGB	TAAAGATGTTGGAGAACTGAAAGA	GACCAGGAAGGACACATACAT
D7S8	870137	C	T	VIC-ctttccCggaatgc-NFQ-MGB	FAM-ctttccTggaaatgc-NFQ-MGB	CTCAGGACCCCTGACCTATTG	GCCTTCTCAGAGCTTATATTGTGC
D10S1257	243	T	C	VIC-tgtgtgttTctgtcc-NFQ-MGB	FAM-tgtgttCtctgtcc-NFQ-MGB	CCCAGGCTGCACAGAAC	TCCTTGACCTAAGACATAGGGC
WIAF1177	686	T	C	VIC-cccAaagaaaca-NFQ-MGB	FAM-cccAaagaaaca-NFQ-MGB	TGGGCTGACTTTCCAA	CAATGCTAATAGACATATGGTAGAT
WIAF1100	667	G	A	VIC-cttacttAcAtctctg-NFQ-MGB	FAM-cttacttAcAtctctg-NFQ-MGB	TGTGGAGCTGAGCACTTT	GACCAAGCTAAACCCAAATACA
WIAF1244	693	T	C	VIC-ccccAttgacatg-NFQ-MGB	FAM-ccccAttgacatc-NFQ-MGB	TCCAACTTCACCTCCATCTCA	GGGCCAGCCTCTTTTCAT
WIAF1296	701	G	A	VIC-actgcttGgccc-NFQ-MGB	FAM-actgcttGgccc-NFQ-MGB	AAAAGAAAGAAATTTAAGGGTGACTTATA	AATGAGGGCTGATCTACCAC
WIAF1430	674	G	A	VIC-ctccagGtcaatgc-NFQ-MGB	FAM-ctccagGtcaatgc-NFQ-MGB	GGTCTCTTTTCTCTAATCTCTAT	TCAAATAATCACTCACCAGTGTGA
WIAF1504	3188	A	G	VIC-tggctTatactgcc-NFQ-MGB	FAM-tggctTatactgcc-NFQ-MGB	CCGCTGATTTTCTTCGACA	CCTAGGCCCTCATCTGTGGAA
WIAF1842	3388	A	G	VIC-ccattTgttacattc-NFQ-MGB	FAM-ccattTgttacattc-NFQ-MGB	CATCTGTTTGTGACATCAGATT	GCAAAGTGCCCATGGTGT
WIAF1634	2043	C	T	VIC-aagattCacaattcc-NFQ-MGB	FAM-aagattCacaattcc-NFQ-MGB	AGGTATTGAACCTTCAATATCTTCAGC	ACATCAATTGAGCTAATGCAAGAA
WIAF1761	3308	G	A	VIC-agcCtgttccct-NFQ-MGB	FAM-agcCtgttccct-NFQ-MGB	GATGGTGAAGCAGCATGAA	CTAGGGTGGCGCTTTAGACA
WIAF2081	1194	T	C	VIC-cacatcaTgtatgaatg-NFQ-MGB	FAM-cacatcaTgtatgaatg-NFQ-MGB	GCAGATCAGTCAGCCACTT	CCCCATCCTACCAGTGGT
WIAF2117	1230	C	T	VIC-cttcaCagtcctct-NFQ-MGB	FAM-cttcaCagtcctct-NFQ-MGB	CCCTTGGCTCTCCACTT	ACCACAGTAGCCATCTGT
WIAF2188	3596	G	A	VIC-aacacaGaaaccagcc-NFQ-MGB	FAM-aacacaGaaaccagcc-NFQ-MGB	CATGCCCAATTTAGATGCT	AGAGCTCAGCGTCTACTCAT
WIAF270	2619	G	A	VIC-atgcccCaagag-NFQ-MGB	FAM-atgcccCaagag-NFQ-MGB	GGCAAAGAACAGAAAGTAAATCCAGAA	CAGTTTCACAGTGAAGTGGCAA
WIAF2072	1185	G	A	VIC-caccattagGtctctg-NFQ-MGB	FAM-caccattagGtctctg-NFQ-MGB	GAAACGCATCTCACTGTCATCTTAT	CACCATACTTCATGCGCAAGACT

Some of TaqMan probes were designed for the antisense sequence (NCBI Assay ID #231, 241, 686, 693, 3188, 3388, 2043, 3308, and 2619). The bases under the Allele 1 and Allele 2 columns were those reported in the dbSNP entries.

Figure 2 Scatter plot of the results of 5'-nuclease assay with marker D2S1301 on 90 individuals. (A) Fluorescence intensity analysis. (B) Normalized fluorescence polarization analysis.

The use of FP as a detection method opens up the possibility of utilizing a somewhat less costly probe (at least 10%–15% cheaper) without the need for the quencher to be present. In addition, FP detection works best when the cleavage reaction is driven to completion and makes it possible to use a limiting amount of the modified probes, thereby reducing the cost of the reaction even further. Whereas the FP plate reader is comparable in cost to the TaqMan plate reader, the throughput is four times higher because it can read 384 reactions in five minutes. If dyes compatible with FP detection can be found across the visible/infrared spectrum, a moderate level of multiplexing (up to 20 alleles, or 10 SNP markers) can be achieved. With the proof-of-principle accomplished, work is under way to design allele-specific probes without the quencher to develop 5'-nuclease assays for FP analysis. One-step, robust assays requiring minimal handling are the methods of choice in DNA diagnostics. The

Table 2. 5'-Nuclease Assay Results for Marker D2S1301 Detected by FP Changes

Sample	Normalized FAM	Normalized VIC	VIC/FAM	GENOTYPE
PD0001	11.2	8.8	0.8	1,2
PD0002	5.2	80.3	15.4	2
PD0003	8.8	80.2	9.1	2
PD0004	8.2	74.8	9.1	2
PD0005	14.8	12.1	0.8	1,2
PD0006	18.3	19	1	1,2
PD0007	17.8	13	0.7	1,2
PD0008	22.7	15.1	0.7	1,2
PD0009	11.8	74.7	6.3	2
PD0010	8.2	73.8	9	2
PD0011	6.8	74	10.9	2
Neg Ctrl	242.6	128.7	0.5	0
PD0012	11.8	79	6.7	2
PD0013	20.8	14.8	0.7	1,2
PD0014	21.4	14.3	0.7	1,2
PD0015	20.5	11.6	0.6	1,2
PD0016	3.8	74.5	19.6	2
PD0017	7.4	74.9	10.1	2
PD0018	6.6	75.7	11.5	2
PD0019	11.5	74.7	6.5	2
PD0020	31.8	19.7	0.6	1,2
PD0021	6.8	74.3	10.9	2
PD0022	82.7	6.3	0.1	1
PD0023	19.2	14.4	0.8	1,2
PD0024	22.4	12.6	0.6	1,2
Neg Ctrl	245	128.5	0.5	0
PD0025	23.8	13.7	0.6	1,2
PD0026	23	13.2	0.6	1,2
PD0027	23.3	13.3	0.6	1,2
PD0028	91.4	5.8	0.1	1

The results of the first 30 samples of the experiment are shown.

Table 3. Concordance Rate of Genotypes Called by FRET or FP Reading of the 5'-Nuclease Assay for 20 Markers

Marker	No. of samples tested	No. of PCR failures	No. of confident calls	Percent concordance
D4S2341	90	0	90	100
D10S1188	90	1	89	100
D4S1301	90	0	90	100
D111964	90	0	88	100
D7S8	90	0	90	100
D10S1257	90	0	90	100
WIAF1177	90	0	90	100
WIAF1100	90	0	90	100
WIAF1244	90	0	90	100
WIAF1296	90	0	90	100
WIAF1430	90	1	89	100
WIAF1504	90	0	90	100
WIAF1842	90	0	90	100
WIAF1634	90	0	90	100
WIAF1761	90	1	89	100
WIAF2081	90	0	90	100
WIAF2117	90	0	90	100
WIAF2188	90	0	89	100
WIAF270	90	0	90	100
WIAF2072	90	0	89	100

FP-TaqMan assay, with reduced setup and operating costs, will make DNA diagnostics more accessible to the average laboratory.

METHODS

Source of DNA

DNA samples from 90 individuals of the NIH Polymorphism Discovery Panel (PD0001-PD0090) were used in this study.

PCR Amplification

Genomic DNA was amplified in a 25 μ L volume using the 5'-nuclease assay kit (PE Biosystems) according to the manufacturer's instructions either on a PE Biosystems 7700 Sequence Detector or on a PE Biosystems 9600 GeneAmp PCR machine. Briefly, the reaction mixture contained 4 ng of genomic DNA, 900 nM each PCR primer, 200 nM each probe, and TaqMan Universal PCR Master Mix (contains AmpErase UNG and AmpliTaq Gold enzymes, dNTPs, and reaction buffer; Applied Biosystems P/N 4304437). TaqMan MGB probes labeled with 6-FAM or VIC (Applied Biosystems P/N 4316033) were used. The reaction mixture was incubated at 50°C for 2 min for AmpErase UNG degradation of any carryover DNA contamination followed by AmpliTaq Gold enzyme activation at 95°C for 10 min before 35 amplification cycles were performed. Each cycle consisted of denaturation at 92°C for 15 sec followed by annealing and primer extension at 60°C for 1 min.

Data Analysis

End-Point Fluorescence Intensity Analysis

The fluorescence intensity of the reaction mixtures was read on the Applied Biosystems 7700 sequence detector using Sequence Detector v1.7 software. The genotypes were called semi-automatically (Livak 1999).

Fluorescence Polarization Analysis

To each reaction well was added 50 μ L of FP reading buffer (50 mM Tris-HCl (pH 9.0), 50 mM KCl, 5 mM MgCl₂, 8% glycerol, 0.1% Triton X-100) and 25 μ L of ethanol. The FP values were determined on a LJI Biosystems (now Molecular Devices) Analyst plate reader. The VIC and FAM dyes were excited at 520 nm and 485 nm and read at 550 nm and 530 nm, respectively, according to the manufacturer's instructions. The FP readings were exported to a spreadsheet and graphed as a scatter plot. The FP readings were normalized by performing two steps: First, the background FP for each dye was subtracted such that the sample with the lowest VIC reading among all samples was given a value of 1 for VIC and that with the lowest FAM reading was given a value of 1 for FAM; next, the scales of both dyes were equalized by multiplying the FAM FP values by a constant such that the average value of the cluster in the upper left corner was the same as that of

the cluster in the lower right corner. The ratios of the normalized VIC and FAM FP values were determined and genotypes were assigned as follows: VIC/FAM <0.2 = allele 1 homozygote; VIC/FAM >5 = allele 2 homozygote; VIC/FAM >0.5 and <3 = heterozygote (if |VIC| and |FAM| FP <15) or negative control/failed PCR (if |VIC| and |FAM| FP >60).

ACKNOWLEDGMENTS

This work is supported in part by a grant from the National Institutes of Health (EY12557) to PYK. S.L. is a recipient of a Howard Hughes Medical Institute summer research fellowship. We thank S. Miller for informatics help.

The publication costs of this article were defrayed in part by payment of page charges. This article must therefore be hereby marked "advertisement" in accordance with 18 USC section 1734 solely to indicate this fact.

REFERENCES

- Afonina, I., Zivarts, M., Kutayavin, I., Lukhtanov, E., Gamper, H., and Meyer, R.B. 1997. Efficient priming of PCR with short oligonucleotides conjugated to a minor groove binder. *Nucleic Acids Res.* **25**: 2657–2660.
- Checovich, W.J., Bolger, R.E., and Burke, T. 1995. Fluorescence polarization: A new tool for cell and molecular biology. *Nature* **375**: 254–256.
- Chen, X., Zehnbaue, B., Gnirke, A., and Kwok, P.-Y. 1997. Fluorescence energy transfer detection as a homogeneous DNA diagnostic method. *Proc. Natl. Acad. Sci.* **94**: 10756–10761.
- Chen, X., Levine, L., and Kwok, P.-Y. 1999. Fluorescence polarization in homogeneous nucleic acid analysis. *Genome Res.* **9**: 492–498.
- Collins F.S., Brooks, L.D., Chakravarti, A. 1998. A DNA polymorphism discovery resource for research on human genetic variation. *Genome Res.* **8**: 1229–1231.
- Kutayavin I.V., Lukhtanov, E.A., Gamper, H.B., and Meyer, R.B. 1997. Oligonucleotides with conjugated dihydropyrroloindole tripeptides: Base composition and backbone effects on hybridization. *Nucleic Acids Res.* **25**: 3718–3723.
- Kwok, P.-Y., Deng, Q., Zakeri, H., and Nickerson, D.A. 1996. Increasing the information content of STS-based genome maps: Identifying polymorphisms in mapped STSs. *Genomics* **31**: 123–126.
- Kwok, P.-Y. 2000. High-throughput genotyping assay approaches. *Pharmacogenomics* **1**: 95–100.
- Lee L.G., Livak, K.J., Mullah, B., Graham, R.J., Vinayak, R.S., Woudenberg, T.M. 1999. Seven-color, homogeneous detection of six PCR products. *Biotechniques* **27**: 342–349.
- Livak K.J. 1999. Allelic discrimination using fluorogenic probes and the 5' nuclease assay. *Genet. Anal.* **14**: 143–149.
- Livak, K.J., Marmaro, J., and Todd, J.A.. 1995. Towards fully automated genome-wide polymorphism screening. *Nature Genet.* **9**: 341–342.

Received November 14, 2000; accepted in revised form December 29, 2000.

Fluorescence Polarization in Homogeneous Nucleic Acid Analysis II: 5'-Nuclease Assay

Sherif Latif, Irma Bauer-Sardina, Kostubh Ranade, et al.

Genome Res. 2001 11: 436-440

Access the most recent version at doi:[10.1101/gr.156601](https://doi.org/10.1101/gr.156601)

References

This article cites 11 articles, 3 of which can be accessed free at:
<http://genome.cshlp.org/content/11/3/436.full.html#ref-list-1>

License

Email Alerting Service

Receive free email alerts when new articles cite this article - sign up in the box at the top right corner of the article or [click here](#).

Affordable, Accurate
Sequencing.

To subscribe to *Genome Research* go to:
<https://genome.cshlp.org/subscriptions>